

# CERTIFICATE OF CALIBRATION



ISSUED BY: SDL ATLAS Ltd.

Date of Issue: 27-Mar-2019

Certificate Number: U19-4033



SDL ATLAS Limited  
1B, Building B,  
Junxiang Da Mansion  
No. 9 Zhongshan Yuan  
Road, Nanshan, Shenzhe  
518057, P.R.C.

## Approved Signatory

Andy Wan	( )
Gary Fenton	( )
Tim Fizer	(√)
Alex Ho	( )

Page 1 of 3

Customer: CERTINTEX SAC  
AV. GRAU 476 SANTA CLARA LIMA, 03  
ATE. VITARTE  
PERU

Instrument: LaunderOmeter (20 pots)

Model: M228AA

Serial Number: 16018/EQ-LAU-01

Methods: The accuracy of this instrument was checked in accordance with:  
BS EN ISO 105 C01 to C05:1993, BS EN ISO 105 C06:2010,  
BS EN ISO 105 C08:2010, BS EN ISO 105 D01 :2010,  
BS EN ISO 105 E03 :2010, AATCC 61,68, 132, 151, 190:2016  
Capacities of the pots were checked using a Measuring Cylinder.  
Linear measurement were checked with Digital Calipers or Steel Ruler  
Temperatures were checked using a Digital Thermometer fitted with a  
Temperature sensor.  
Speed was checked using a Stopwatch.

Date of Calibration: 27-Mar-2019

Signature.....

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.